

CERTIFICATE OF ANALYSIS

ACIRS-H8-2021-Lot #1

Certified Reference Materials for

Hardgrove Grindability Index

(Set of 4)

Certified:	August 2021
Valid to:	November 2022
Report Number:	ACIRS-H8-Lot#1-CoA-01
Previous ACIRS-H series:	Supersedes ACIRS-H7-2019

1. Introduction

ACIRS-H8-2021-Lot#1 is a Certified Reference Material comprising a set of four jars each having a different Hardgrove Grindability Index (HGI) value. Samples have a nominal mass of 1 kg and top-size of 4.75 mm.

Sample preparation, homogeneity assessment and certification have been conducted by an AS ISO/IEC 17025 accredited facility in accordance with ASTM D409/D409M-16.

Production and certification was conducted in accordance with the technical and production requirements of ISO 17034.

The intended use of these samples is as a quality control tool and for calibration of Hardgrove grindability machines.

2. Certified Values

Table 1 ACIRS-H8-2021-Lot#1 Certified Values*

ACIRS-H8-2021 Lot#1	Hardgrove Grindability Index ^a (HGI)	Standard Deviation ^b	Uncertainty ^c (k=2)	No. of Samples Tested	Repeatability
Sample A	27	0.2	0.1	19	3 units
Sample B	44	0.2	0.1	19	3 units
Sample C	59	0.3	0.1	19	3 units
Sample D	83	0.4	0.2	19	3 units

* This is an empirical method. All values are provided in HGI units which have no absolute value. Certified values are valid when tested according to ASTM D409/D409M-16 and equivalent methods. The grindability characteristics of samples may be altered by conditions during drying and preparation. Table 3 reflects average sample moisture and relative humidity at the time of testing.

NOTES

- HGI property values are the best estimate of the true HGI value and are based on the unweighted mean of means. Characterisation was conducted by ASTM D409/D409M-16 with a primary certified reference material set supplied by Penn State University using the Australian national Hardgrove machine.
- Standard deviation (sd) is the standard deviation of sample means tested under repeatability conditions.
- Uncertainty is an expanded uncertainty and provides the user with information on the likely range of the true (but unknown) HGI value and has been estimated in accordance with the Guide to the Expression of Uncertainty in Measurement (GUM) with a coverage factor k=2, corresponding to a level of confidence of about 95%. It has been derived from the observed standard deviation of the population mean which includes contribution from sample inhomogeneity. The contribution from instability is considered negligible.

3. Traceability

Empirical HGI values for ACIRS-H8-2021-Lot#1 are traceable to the primary certified reference materials produced by Penn State University, as listed in Table 2, when analysed by ASTM D409/D409M and equivalent methods.

4. Instructions for Handling and Use

Sample bottles should be kept tightly sealed and stored in a cool, dark place. Do not freeze.

The reference material should be thoroughly mixed by end-over-end rotation before sub-sampling. Samples should be prepared and analysed in accordance with the most recent version of ASTM D409/D409M or equivalent test methods (as above).

Minimum sample size is in accordance with ASTM D409/D409M or equivalent test methods.

The Safety Data Sheet is available from www.acirs.com.au/products/hardgrove-grindability/

5. Source and Preparation

Samples of mass greater than 250 kg of each of 4 coals were obtained:

Sample A: High volatile thermal coal, South-East Qld

Sample B: High volatile thermal coal, Hunter Valley, NSW

Sample C: High rank bituminous thermal coal, Central Qld

Sample D: High rank bituminous coking coal, Central Qld

192 x 1kg sub-samples for each of A, B, C and D were prepared in strict accordance with Annex A1-A2 of ASTM D409/D409M. Blending was conducted by multiple rotary sample division steps.

6. Homogeneity Assessment and Certification

Confirmation of satisfactory homogeneity was conducted in strict accordance with Annex A3 of ASTM D409/D409M. The Australian national HGI machine was calibrated against primary certified reference materials as shown in Table 2. This data was used in the creation of the calibration graph used to certify for ACIRS-H8-2021¹.

Table 2 Calibration of National Hardgrove Machine

ASTM certified reference material set			
Serial number	HGI (units)	Mean mass - 75 µm (g)	Repeatability (units)
2021-40-04 (A)	42	4.16	3
2021-40-05R* (B)	59	6.96	3
2021-40-04 (C)	68	8.34	3
2021-40-04 (D)	92	11.31	3
Linear regression HGI = 6.982x + 11.532 (R ² = 0.993)			

*Sample B from the 2021-40-04 set was recalled by Penn State University and replaced by 2021-40-05R

19 samples² were selected from each of Lots A, B, C and D by a process of random systematic sampling and analysed in duplicate against this calibration line.

The HGI values and standard deviation of each sample is provided in Table 3. Samples A, B, C and D met the criteria for satisfactory homogeneity as specified in Annex A3 of ASTM D409/D409M-12. ACIRS-H8-2021 therefore passed homogeneity test criteria.

¹ Continuity of calibration was confirmed by comparing calibration lines used for certification of ACIRS- H7-2019 (Lot #2).

² This represents 10% of the total production size for ACIRS-H8-2021

Table 3 ACIRS-H8-2021 homogeneity and certification data*

	SAMPLE A		SAMPLE B		SAMPLE C		SAMPLE D	
	-75 µm mass (g)	HGI	-75 µm mass (g)	HGI	-75 µm mass (g)	HGI	-75 µm mass (g)	HGI
Average	2.22	27.0	4.64	43.9	6.85	59.3	10.23	83.0
Standard Deviation	0.02	0.17	0.03	0.23	0.04	0.28	0.06	0.41
No. samples	19		19		19		19	
% Yield, -1.18 x 0.6 mm	67.1		67.1		67.1		62.8	
Environmental conditions during testing								
Relative Humidity (%)	Mean: 47 Range: 46 - 50		Mean: 37 Range: 35 - 38		Mean: 46 Range: 44 - 48		Mean: 52 Range: 51 - 55	
Air dried moisture (%)	Mean: 4.12 Range: 4.00 – 4.20		Mean: 3.13 Range: 3.07 – 3.18		Mean: 2.51 Range: 2.41 – 2.57		Mean: 1.11 Range: 1.09 – 1.14	

* Based on Regression Equation in Table 2

The analysis conducted for the determination of satisfactory homogeneity is equivalent to certification of Lot #1. Certification data has been provided in Table 1.

Revision History

Document Number	Summary	Date
ACIRS-H8-Lot#1-CoA-01	Original	23/08/2021

Disclaimer

To the extent permitted by law, ACIRS disclaims all warranties whether expressed or implied with regard to merchantability, non-infringement, fitness for a particular purpose. In no event will ACIRS be liable for incidental damage or consequential loss arising from the use of this product. Where the product does not conform to the certified values, giving due consideration to the stated uncertainties and accepted tolerances, the total liability of ACIRS shall be limited at ACIRS' absolute discretion to either replacement of the product or refund of the purchase price.

Approved by: Christine Foster (ACIRS Operations Manager)
Australian Coal Industry Reference Samples (ACIRS)

PO Box 2315, DANGAR NSW 2309, AUSTRALIA

Phone +61 (2) 4926 4870

Fax +61 (2) 4926 4902

Email acirs@acps.com.au