

CERTIFICATE OF ANALYSIS

ACIRS-S1D-2011

Reference Material for determination of total sulfur in coal containing a minimum mass of 250 g at a nominal top-size of -212 μm .

ACIRS-S1D-2011	Units	Total Sulfur	Standard Deviation
	%, dry basis	2.47	0.043

* Certified against AS 1038.6.3.3 for sulfur and AS1038.3 for moisture.

Samples should be thoroughly mixed by end-over-end rotation before sub-sampling. Sample bottles should be kept tightly sealed and stored in a cool, dark place.

Date of Certification: December, 2011
Valid to: December, 2018
Approved by: Mark Bennetts

The full technical report and SDS for this product are available at
www.acps.com.au/acirs/

Australian Coal Industry Reference Samples (ACIRS)
PO Box 2315, DANGAR NSW 2309 AUSTRALIA

Phone: +61 (2) 4926 4870

Fax: +61 (2) 4926 4902

Email: acpsnational@acps.com.au